



## ಬೃಹತ್ ಬೆಂಗಳೂರು ಮಹಾನಗರ ಪಾಲಿಕೆ

ಕಾರ್ಯಪಾಲಕ ಅಭಿಯಂತರರು, ಮಲ್ಲೇಶ್ವರಂ ವಿಭಾಗ, 2ನೇ ಅಡ್ಡರಸ್ತೆ, ಜಲದರ್ಶಿನಿ ಲೇಔಟ್, ಎಂ.ಎಸ್.ಆರ್. ನಗರ, ಮತ್ತಿಕೆರೆ, ಬೆಂಗಳೂರು -560054.

ಸಂಖ್ಯೆ: ಕಾ.ಪಾ.ಅ(ಮ.ವಿ)/ಕೋ.ಪ್ರ/08/ 2020-21

ದಿನಾಂಕ:-23-12-2020

### ಕೋಟೆಷನ್ ಪ್ರಕಟಣೆ


ವಿಷಯ: ಬೃಹತ್ ಬೆಂಗಳೂರು ಮಹಾನಗರ ಪಾಲಿಕೆ ವ್ಯಾಪ್ತಿಯ ಮಲ್ಲೇಶ್ವರಂ ವಿಭಾಗ ಸಭಾ ಕ್ಷೇತ್ರದ ನಗರ ಪ್ರಾಥಮಿಕ ಆರೋಗ್ಯ ಕೇಂದ್ರಗಳಲ್ಲಿ ಕೋವಿಡ್‌ನಿಂದ ಗುಣಮುಖರಾದ ಸಾರ್ವಜನಿಕರಿಗೆ ಹೆಚ್ಚಿನ ಪರೀಕ್ಷೆ ನಡೆಸಲು ಅಗತ್ಯವಿರುವ ಲ್ಯಾಬ್ ರಿಪಜೆಂಟ್ಸ್ ಮತ್ತು ಉಪಕರಣಗಳನ್ನು ಖರೀದಿಸಲು ಕೊಟೆಷನ್ ಕರೆಯುವ ಬಗ್ಗೆ.

ಮೇಲ್ಕಂಡ ವಿಷಯಕ್ಕೆ ಸಂಬಂಧಿಸಿದಂತೆ ಬೃಹತ್ ಬೆಂಗಳೂರು ಮಹಾನಗರ ಪಾಲಿಕೆ, ಮಲ್ಲೇಶ್ವರಂ ವಿಭಾಗ ವ್ಯಾಪ್ತಿಯಲ್ಲಿ ಕೋವಿಡ್‌ನಿಂದ ಗುಣಮುಖರಾದ ಸಾರ್ವಜನಿಕರಿಗೆ ಹೆಚ್ಚಿನ ಪರೀಕ್ಷೆ ನಡೆಸಲು ಅಗತ್ಯವಿರುವ ಲ್ಯಾಬ್ ರಿಪಜೆಂಟ್ಸ್ ಮತ್ತು ಉಪಕರಣಗಳನ್ನು ಖರೀದಿಸಲು ಉದ್ದೇಶಿಸಿದ್ದು, ಅರ್ಹ diagnostics equipments ಸರಬರಾಜುದಾರರುಗಳಿಂದ ಸರಬರಾಜು ಮಾಡಿಕೊಳ್ಳಲು ಉದ್ದೇಶಿಸಿದ್ದು, ಇಚ್ಛೆಯುಳ್ಳವರು ತಮ್ಮ ಕನಿಷ್ಠ ದರಗಳನ್ನು ದಿನಾಂಕ:-29-12-2020 ರಂದು ಸಂಜೆ 4.00 ಗಂಟೆಯೊಳಗೆ ಈ ಕೆಳಗೆ ಸಹಿ ಮಾಡಿದವರಿಗೆ ಖುದ್ದಾಗಿ ಸೀಲ್ ಮಾಡಿದ ಲಕೋಟೆಯಲ್ಲಿ ತಮ್ಮ ಕನಿಷ್ಠ ದರಗಳನ್ನು ಸಲ್ಲಿಸತಕ್ಕದ್ದು. ತದ ನಂತರ ಬಂದ ಕೊಟೆಷನ್‌ಗಳನ್ನು ಸ್ವೀಕರಿಸಲಾಗುವುದಿಲ್ಲ ಹಾಗೂ ಅದೇ ದಿನ ಸಂಜೆ 4.30 ಗಂಟೆಗೆ ಸಲ್ಲಿಕೆಯಾದ ಕೊಟೆಷನ್‌ಗಳನ್ನು ತೆರೆಯಲಾಗುವುದು.

### Procurement of lab equipments for Health centers

SI No	Qty	Particulars	Unit
1	1	Fixation of rate for Procurement of Discrete random access analyser system;180 test/hour constant;Linear and nonlinear.Logit-4P,Logit-5P,Spline,exponent,polynomial,factor method of Calibration Curve;30 Position;2-30QuL;Reagent(Postion:60;Volume:20-30QuL;Cooling Temperature:4-16°C);Reaction(ReactionCuvette:81 semi disposal plastic cuvettes;Reaction Temperature:37 °C ±0.2 °C ;optical Source:Tungsten Halogen Lamp; Photometry:8 filters; Photometry Linearity:0-3.8Abs);Operating system-Windows 10;Barcode reading Reagent & Sample available;TURBOCHEM PRIME Automated Random Access Biochemistry Analyzer and other allied equipments for Malleshwaram Division	Nos

2	1	Fixation of rate for Procurement of 60 Samples/hour;25 Routine test;4 Reasearch Parameters;Graphics:2histograms for RBC & PLT,& 4 Scattergrams for WBC Differential;Sample Volume Whole Blood-20µL, Prediuted Blood-20µL;Analyser Mode-CBC,CBC+Diff, 10.4 inch Touch screen LCD NosDisplay;3 Level QC,LJ,XB of quality Cortrol;Foot print size-430mm(L)*350mm(W)*435mm(H) PENTACYTE-5 Part differential Hematology Analyzer and other allied equipments Mallechwaram Division	Nos
3	1	Fixation of rate for Procurement of Test Speed-20min/cycle;Model Type-8/16/32 sample Positions;Mode-Cycle/Random;Display-7inchs LCD touch Screen;Internal Barcode Reader;ESR Dynamic Curve;Internal QC Management;Temperature Correction;Internal Mixing TurboLyte-Electrolyte Analyzer and other allied equipments Mallechwaram Division	Nos

  
 ಕಾರ್ಯಪಾಲಕ ಅಭಿಯಂತರರು  
 ಮಲ್ಲೇಶ್ವರಂ ವಿಭಾಗ  
 ಬೃಹತ್ ಬೆಂಗಳೂರು ಮಹಾನಗರ ಪಾಲಿಕೆ  
